



MSI-B 2020 Microsatellite Instability

INSTITUTION: Genetiks Genetic Diseases Diagnosis Center
Istanbul TU 34365

ATTENTION: Izzet Akiva

CAP NUMBER: 8732280-01 **Kit#** 1

KIT INFORMATION:	Kit ID: 33771850	Kit Mailed: 10/19/2020	Original Evaluation: 12/21/2020	Next Mailing Date: 6/14/2021
-------------------------	----------------------------	----------------------------------	---	--

COPIED TO: CAP

LEGEND: Exception Reason Codes appearing in this evaluation:
<NONE>

Reviewed By

Date



**COLLEGE of AMERICAN
PATHOLOGISTS**

325 Waukegan Road, Northfield, Illinois 60093-2750
800-323-4040 • cap.org

CAP Number: 8732280-01

Kit# 1

Kit ID: 33771850

Institution: Genetiks Genetic Diseases Diagnosis Center

Kit Mailed: 10/19/2020

Attention: Izzet Akiva

Original Evaluation: 12/21/2020

City / State: Istanbul TU 34365

**EVALUATION
ORIGINAL**

MSI-B 2020 Microsatellite Instability

Test Method	Specimen	Your Result	Intended Response	Your Grade
MSI Result	MSI-04	MSS	MSS	Good
MLH1 Result	MSI-04	TEST NT PERF TISSUE MSS	NO MLH1 PROM HYPERMETH TEST NT PERF TISSUE MSS	Good
MSI PP1 Result	MSI-05	MSS	MSS	Good
MSI PP2 Result	MSI-06	MSI-H	MSI-H	Good

The College of American Pathologists recommends that the result of this interlaboratory comparison not be used as a sole criterion for judging the performance of any individual clinical laboratory.